

SREMS Minimum Equipment Standards

Standard Description	First Response Vehicle							
	Ambulance		AEMT		EMT-CC		EMT-P	
	Opt	Req	Opt	Req	Opt	Req	Opt	Req
Unit meets NYS minimum BLS equipment standards for certification, per latest Part 800.		X						
Portable monitor-defibrillator with the following:		1				1		1
Operates on batteries, independent of outside power source		X				X		X
Monitors patient's Lead II ECG via self-adhesive defibrillation electrodes		X				X		X
Cardiac monitoring in operator's manually-selectable choice of standard limb leads I, II, and III, via a standard three- or four-electrode patient monitoring cable and standard (dot) monitoring electrodes		X				X		X
Continuous and/or intermittent real-time recording of the patient's ECG on paper + highlight replay (Code Summary)		X				X		X
Manual defibrillator AED mode		X				X		X
Self-adhesive defibrillation electrodes		X				X		X
Manually-actuated, synchronized cardioversion		X				X		X
Interconnectivity with biotelemetry transmitter		X				X		X
External transcutaneous pacemaker		X				X		X
Make & model of monitor defibrillator:								
Month & year purchased:								
Purchased: <input type="checkbox"/> New <input type="checkbox"/> Used								
Sets of adult-sized, self-adhesive, disposable defibrillation electrodes OR sets of defibrillation pads (for hand-held paddles) OR tubes/bottles of defibrillation paste or gel		2						
Sets of pediatric-sized self-adhesive defibrillation electrodes		2						
Sets of self-adhesive, disposable external pacing electrodes (may be same as ALS 2, if unit uses combination electrodes)		2						
Spare monitoring cable for monitor/defibrillator (not required if monitor has hands-free defib)		1						
Cardiac Monitor/Defibrillator						1		1
AED				1				
ECG monitoring electrodes - 12-lead acquisition		1						
12-Lead Transmission						1		1
Continuos Waveform Capnography		1		1		1		1
Pulse Oximeter		1		1		1		1
CPAP	BLS	ALS		1		1		1
Glucometer with test strips & lancets	BLS			1		1		1
Nebulizer		3		3		3		3
Atomizer		2		2		2		2
End Tidal CO2 Detector-Adult				1		1		1
End Tidal CO2 Detector-Pediatric						1		1

SREMS Minimum Equipment Standards

Standard Description	First Response Vehicle							
	Ambulance		AEMT		EMT-CC		EMT-P	
	Opt	Req	Opt	Req	Opt	Req	Opt	Req
Esophageal Detector Device						1		1
Surgical Airway Device								1
Arterial Tourniquet				2		2		2
Gum Elastic Boogie			1			1		1
Length-Based Resuscitation Tape						1		1
Endotracheal Tubes								
(ID uncuffed?) 2.5 mm		2						2
(ID uncuffed?) 3.0 mm		2						2
(ID uncuffed?) 3.5 mm		2						
(ID uncuffed?) 4.0 mm		2						2
(ID uncuffed?) 5.0 mm		2						2
(ID cuffed?) 6.0 mm		2		2		2		2
(ID cuffed?) 7.0 mm		2		2		2		2
(ID cuffed?) 8.0 mm		2		2		2		2
(ID cuffed?) 9.0 mm		2		2		2		2
Commercial ET Tube Securing Devices		2		2		2		2
Water Soluble Jelly								
Tube or Packets				1		1		1
				4		4		4
10 cc Syringe for Endotracheal Tube				2		2		2
Regionally Approved Alternative Airway Device	√							
>Adult Set				1		1		1
>Pediatric Set								1
Adult Stylet (can be packaged with ETT)		2		2		2		2
Pediatric Stylet (can be packaged with ETT)		2				2		2
Infant-sized Stylet		2						
Laryngoscope Handle				1		1		1
Laryngoscope Handle with "C" Batteries		1						
Laryngoscope Handle with "AA" Batteries		1						
Laryngoscope Blades								
Curved-2				1		1		1
Curved-3				1		1		1
Curved-4				1		1		1
Straight-0								1
Straight-1								1
Straight-2				1		1		1
Straight-3				1		1		1
Straight-4				1		1		1
Miller #0 Lighted Laryngoscope Blade		1						
Miller #1 Lighted Laryngoscope Blade		1						
Miller or Mac #2 Lighted Laryngoscope Blade		1						
Miller or Mac #3 Lighted Laryngoscope Blade		1						
Miller or Mac #4 Lighted Laryngoscope Blade		1						
Adult Magill Forceps		2		1		1		1
Pediatric Magill Forceps		2				1		1

SREMS Minimum Equipment Standards

Standard Description	First Response Vehicle							
	Ambulance		AEMT		EMT-CC		EMT-P	
	Opt	Req	Opt	Req	Opt	Req	Opt	Req
Suction Catheters				2		2		2
8 French				2		2		2
10 French				2		2		2
12 French Yankauer				2		2		2
14 Gauge x3" IV Catheter						2		2
Orogastric Tubes								
14 French								1
16 French								1
1000 mL Normal Saline				2		2		2
100 mL Normal Saline						2		2
Macro drip IV Admin. Set				2		2		2
Micro drip IV Admin. Set						2		2
Large Bore Saline Lock Kit				4		4		4
Protective IV Catheter 14G x 1 1/4"				2		2		2
Protective IV Catheter 16G				2		2		2
Protective IV Catheter 18G				2		2		2
Protective IV Catheter 20G				2		2		2
Protective IV Catheter 22G						2		2
Protective IV Catheter 24G						2		2
#14 gauge over-the-needle IV catheters		3						
#16 gauge over-the-needle IV catheters		3						
#18 gauge over-the-needle IV catheters		3						
#20 gauge over-the-needle IV catheters		3						
#22 gauge over-the-needle IV catheters		3						
#24 gauge over-the-needle IV catheters		3						
#18 gauge steel hypodermic needles		2						
#21 gauge steel hypodermic needles		2						
#25 gauge steel hypodermic needles		2						
Bone Marrow Needles		2						
20 mL Luer-lock sterile disposable syringe		2						
10 mL Luer-lock sterile disposable syringe		2						
5 mL Luer-lock sterile disposable syringe		2						
3 mL Luer-lock sterile disposable syringe		2						
1 mL Luer-lock sterile disposable syringe		2						
Intraosseous Needle - Adult (powered device)				1		1		1
Intraosseous Needle - Peds				1		1		1
1cc Syringe with 25G Needle				2		2		2
3cc Syringe						2		2
5cc Syringe						2		2
10cc Syringe						2		2
18G Hypodermic Needle (If packaged with a syringe, this will count toward the minimum quantity.)						2		2
22G Hypodermic Needle (If packaged with a syringe, this will count toward the minimum quantity.)						2		2

SREMS Minimum Equipment Standards

Standard Description	First Response Vehicle							
	Ambulance		AEMT		EMT-CC		EMT-P	
	Opt	Req	Opt	Req	Opt	Req	Opt	Req
10 mL Disposable Syringes without Needles		2						
Disposable Razors				1		1		1
Disposable IV Tourniquets (If packaged with a saline lock kit, this will count toward the minimum quantity.)				2		2		2
Alcohol Prep Pads (If packaged with a saline lock kit, this will count toward the minimum quantity.)				12		12		12
Transparent Dressings (If packaged with a saline lock kit, this will count toward the minimum quantity.)				2		2		2
2" x 2" Gauze Pads (If packaged with a saline lock kit, this will count toward the minimum quantity.)				12		12		12
Iodine Pads or Chlorhexidine Wipes								3
Band-aids				12		12		12
9" Padded IV Armboard				1		1		1
Spare Patient Monitor Cables						1		1
EKG Electrode Set - Adult						2		2
EKG Electrode Set - Pediatric						2		2
AED Pad - Adult **				2				
AED Pad - Pediatric **				2				
Pacing/Defibrillation Pad - Adult						2		2
Pacing/Defibrillation Pad - Pediatric						2		2
Rolls 1/2" or 1" Cloth Adhesive Tape		2						
Broselow Pediatric Emergency Tape		1						
Box Small Non-sterile Exam Gloves		1						
Box Medium Non-sterile Exam Gloves		1						
Box Large Non-sterile Exam Gloves		1						
Pairs of Splash-protective-type goggles/faceshields		2						
Disposable Surgical Face Masks OR		12						
N-95 or Greater Respirator for EACH Crew Member on Vehicle		2						
Disposable, Fluid-Impermeable Medical Gowns		2						
Mechanical IO Gun/Drill/Spring Loaded Device	√							
100 mL Plastic Bags of Normal Saline .09%		6						
1000 mL Plastic Bags of Normal Saline		4						
MEDICATIONS								
Adenosine 6 mg/2 mL		3				6		6
Albuterol 2.5 mg/3 mL				4		4		4
Albuterol Premix 0.083% (3 mL)		5						
Amiodarone, vial, 150 mg (3 mL)		4				4		4
Aspirin, bottle, chewable 325 or 81 mg/tab		4		1		1		1
Atropine Sulfate 1 mg/10 mL Syringe						2		2
Atropine Sulfate 1.0 mg/10 mL		2						
Benadryl 50 mg/1 mL (Diphenhydramine)		1						
Calcium Chloride 100 mg/1 mL		2						2
Dexamethasone (Decadron) 10 mg/1 mL		2				2		2
Dextrose 10% 25 grams/unit 250 mL		2		2		2		2

SREMS Minimum Equipment Standards

Standard Description	First Response Vehicle							
	Ambulance		AEMT		EMT-CC		EMT-P	
	Opt	Req	Opt	Req	Opt	Req	Opt	Req
Diltiazem 5 mg/1 mL						2		2
Diltiazem 25 mg/5 mL, may sub Lopressor		1						
Epi-Pen Autoinjector - Adult				1				
Epi-Pen Autoinjector - Pediatric				1				
Epinephrine 1:1,000 1 mg/1 mL (Ampule)		2				3		3
Epinephrine 1:10,000 1 mg/1 mL				6		6		6
Epinephrine 1:10,000 0.1 mg/10 mL		6						
Etomidate 2 mg/mL if authorized		2						
Etomidate 20 mg/10 mL Vial						3		3
Fentanyl 100 mcg/2 mL Ampule						2		2
Fentanyl 100 mcg		2						
Glugagon 1 mg/1 mL Vial				1		1		1
Glucagon 1 mg unit		1						
Glucose - Oral 31 gm				1		1		1
Glucose Paste (tube)		1						
Haloperidol 5 mg/1 mL Ampule						2		2
Haloperidol		2						
Ipratropium Bromide 500 up/2.5 mL or Ipratropium & Albuterol						3		3
Ipratropium (Atrovent)-May be premixed with albuterol		2						
Ketamine 100 mg/1 mL								2
Ketamine (paramedic only) 500 mg/5 mL		2						
Ketorolac 30 mg/1 mL Vial						2		2
Ketorolac (Toradol)		1						
Lidocaine 100 mg/5 mL						3		3
Lidocaine 2%		3						
Magnesium Sulfate (Total 8 gm per vehicle)						1		1
Magnesium Sulfate 5 gm		2						
Metoprolol 5 mg/5 mL Ampule						1		1
Metoprolol		4						
Midazolam 5 mg/5 mL						4		4
Midazolam (Versed) 5 mg		4						
Morphine Sulfate 10 mg/1 mL						2		2
Morphine 10 mg		2						
Naloxone 2 mg/2 mL				2		2		2
Naloxone (Narcan) 1 mg		2						
Nitroglycerin 0.4 mg Tablet (bottle) and/or Nitroglycerin Spray						1		1
Nitroglycerin 0.4 mg/tab, 25 tabs/bottle or Metered Dose		1						
Nitrous Oxide 50/50	√							
Norepinephrine 4 mg/4 mL						1		1
Norepinephrine 4 mg		1						
Ondansetron 4 mg/2 mL Vial						3		3

SREMS Minimum Equipment Standards

Standard Description	First Response Vehicle							
	Ambulance		AEMT		EMT-CC		EMT-P	
	Opt	Req	Opt	Req	Opt	Req	Opt	Req
Ondansetron 4 mg Tablets or 4 mg Orally Disintegrating Tabs (ODT) 1 Bottle or Unit Dose Packs						1		1
Ondansetron (inj) 4 mg		2						
Ondansetron (PO) 4 mg tab		2						
Rocuronium 50 mg/5 mL Vial								2
Rocuronium 10 mg (RSI Med Optional)	2							
Sodium Bicarbonate 44.6 meq Syringe								2
Sodium Bicarbonate 50 meq/50 mL		2						
Succinylcholine 100 mg/5 mL								2
Succinylcholine 20 mg (RSI Med Optional)	2							
Tetracaine Unit Dose	√							
Vecuronium 10 mg/10 mL Vial								2
Vecuronium 1 mg (RSI Med Optional)	√							
Mobile Vehicle Radio, at least 25 watts, with: Transmit/receive on 155.340 MHz Transmit/receive on 155.400 MHz		1						
		√						
		√						
Copy of Current Regional ALS Protocols (electronic is acceptable)		1		1		1		1
All ALS equipment and supplies clean, sanitary, and/or sterile (where appropriate)		√						
Drugs, sharps and syringes securely locked in fixed cabinet(s) in vehicle		√						

** - AED PAD- Adult/Pediatric- may be combined if device supports pediatric levels in adult pad.